





Alterra 1130 Hi / Low Long-Term Care Hospital Beds

Alterra 4-Motor Full-Electric High-Low Hospital Bed, 35" Wide, Bed Only Maximum height of 26" for easy resident

access; low height of 7.25" for Alzheimer's patients and any other resident at risk of falling

- · Smooth performance with four ultra-quiet, dependable DC motors: two for height, one for knee hinge, and one for reclining head deck
- Grid and rib combination deck design provides comfort and strength
 Bed deck dimensions: 80" L x 34.5" W; 87.5" overall length
- Battery backup not included

Packaging	1 Each / Each	Report Incorrect Product Information
Manufacturer	Medline	
Manufacturer #	FCE1130B	



Account Linked Item Last Ordered on 06/15/2023 View Order History

Availability Details

Shipping from Saint Peters, MO (C37). While the above is currently available, it may differ at time of order.

Contract & Account Information

Customer Number 0001996428 Customer Name SKY MEDICAL SUPPLIES LLC Contract Item Yes **GPO** Information Mfr. Contract # MEDLINE MEDLINE Medline Contract # Contract Name **Special Pricing Agreement** Tier N/A 01/01/2023 Valid From Valid To 12/31/2023

Specifications

BATTERY BACKUP No INCLUDED BED STYLE Hi-Low BRAKE AT FOOT No SECTION INCLUDED **DECK STYLE** Grid & Rib Deck LATEX FREE UNSPSC 42191807

Unit of Measure	Conversion	Net/Gross Weight (lbs)	Volume (cubic ft)	Shipping Dimensions (inch) L x W x H	GTIN
Each (EA)	1.0 Each	0.0 / 231.483	14.719	36.5 x 15.5 x 48.5	10888277694521

HCPCS Information

HCPCS Code: NO GOVERNMENT OR SUPPLIER HCPCS CODE VERIFICATION

HCPCS Disclaimer: HCPCS codes and Home Health Consolidated Billing codes provided by Medline are intended as general guidelines only. Medline does not guarantee coverage or reimbursement of any products. You must address all coverage and reimbursement issues (including the correctness and accuracy of codes) with your individual payers. It is your responsibility to ensure the accuracy and appropriateness of each claim you submit, in accordance with all applicable payer requirements.